

## EC Declaration of Conformity

**Manufacturer:**

Shenzhen Comen Medical Instruments Co.,Ltd  
**Address:**

Floor 10, Floor 11 and Section C of Floor 12 of  
 Building 1A & Floor 1 to Floor 5 of Building 2,  
 FIYTA Timepiece Building, Nanhuan Avenue,  
 Matian Sub-district, Guangming District,  
 Shenzhen, Guangdong, 518106, P.R. China.

**Whose Single Authorized Representative:**

Lotus NL B.V.

**Address:**

KoninginJulianaplein 10, 1e Verd, 2595AA,  
 The Hague, Netherlands.

We, the manufacturer, declare at our sole responsibility that following products

Product name	Model
Syringe Pump	M300, M500

meet the provisions of Directive 93/42/EEC.

The medical device has been assigned to class IIb according to rule 11 in Annex IX of the Directive 93/42/EEC. It bears the mark

CE 1639

The product concerned has been designed and manufactured under a quality management system according to Annex II (excluding Section 4) of Directive 93/42/EEC.

The product meet the following standard: (See Chapter 3 of File No. 1206-089-01)

Compliance of the designated product with the Annex II (excluding Section 4) of Directive 93/42/EEC has been assessed and certified by the Notified Body

**SGS Belgium NV**

**SGS House Noorderlaan  
 87 2030 Antwerp Belgium**

CertificateNo.: CN19/41057

Issuedate: 2021.03.22

Expirydate: 2023.02.05

The above mentioned declaration of conformity is exclusively under the responsibility of

Company: Shenzhen Comen Medical Instruments Co.,Ltd  
 Address: Floor 10, Floor 11 and Section C of Floor 12 of Building 1A & Floor 1  
 to Floor 5 of Building 2, FIYTA Timepiece Building, Nanhuan Avenue,  
 Matian Sub-district, Guangming District, Shenzhen, Guangdong,  
 518106, P.R. China.

Shenzhen, 2021.05.08  
 Place, date

Comghun Management Representative  
 Legally binding signature, Function

